Technical Note

The cleaners listed below are tested for use on Datalogic's Standard and Healthcare Enclosures. (all the Gryphon 4500 series).



PRODUCT	CHEMICAL CONTENT
Alcohol Wipes	70% Isopropyl Alcohol
Hydrogen Peroxide	3%
Formula 409® Glass and Surface Cleaner	n-Alkyl Dimethyl Benzyl Ammonium Chloride; n-Propoxypropanol
Windex® Blue	2-hexoxyethanol, isopropanolamine, sodium dodecylbenzene sulfonate, Lauryl Dimethyl Amine Oxide, ammonium hydroxide
100% Gentle dish soap and water	

NOTE: Typical rubbing alcohol usually has the correct mixture of 70% isopropyl alcohol / 30% water. No additional water needed. (Check label to be sure)

The cleaners and disinfectants listed below are tested for use on Datalogic's Disinfectant-Ready Enclosures (Gryphon 4500 Healthcare series only).

- REFER TO THE DEVICE INSTRUCTION MANUAL FOR PROPER CLEANING METHODS WHEN USING THESE CHEMICALS.
- DISINFECTANTS SHOULD NOT BE SPRAYED OR POURED ONTO THE SCANNER UNIT.
- THE SCANNERS MUST ALWAYS BE WIPED DRY AFTER DISINFECTING.





PRODUCT	CHEMICAL CONTENT
Sani-Cloth® HB, Sani-Cloth® Plus, Super Sani-Cloth®	Quaternary Ammonium Chloride solution
Hepacide Quat II	Virucidal disinfectant cleaner
Alcohol Wipes	70% Isopropyl Alcohol
CaviWipes™	Isopropanol 10-20%; Ethylene Glycol Monobutyl Ether 1-5%
Virex® 256	n-Alkyl Dimethyl Benzyl Ammonium Chloride; Didecyl Dimethyl Ammonium Chloride
Formula 409® Glass and Surface Cleaner	n-Alkyl Dimethyl Benzyl Ammonium Chloride; n-Propoxypropanol
Windex® Blue	2-hexoxyethanol, isopropanolamine, sodium dodecylbenzene sulfonate, Lauryl Dimethyl Amine Oxide, ammonium hydroxide
Clorox® Bleach; Clorox Healthcare Bleach Germicidal Cleaner	Sodium Hypochlorite; Sodium Hydroxide
Hydrogen Peroxide	3%
ProSpray™ Wipes (Disinfectant towelettes)	0.647% o-phenylphenol; 0.070% o-benzyl-p-chlorophenol
100% Gentle dish soap and water	